

Routine Determination of Mirex and Photomirex in Fish  
Tissue in the Presence of Polychlorinated Biphenyls

A Thesis

Presented to the Faculty of the Department of Biological Sciences  
of the State University of New York College at Brockport  
in Partial Fulfillment for the Degree of  
Master of Science

by

Samuel E. Insalaco

December, 1979

### Acknowledgements

I am deeply indebted to my advisory committee, Dr. Joseph C. Makarewicz, Dr. Martha M. Vestling, and Dr. James M. Haynes. I thank Joe for his wisely guidance, understanding and subtle motivation. I thank Marty for her undying encouragement and for expanding my awareness into adjacent fields. I thank Jim for his tremendous enthusiasm toward my work.

I would also like to thank Dr. Theodore J. Starr and the Department of Biological Sciences, Sandra Bauch, and Terace Rohland who has been more than a friend. I would also like to thank all my other friends in Brockport, especially those in the Chemistry Department.

## Dedication

This thesis is dedicated primarily to my mother and father, Madeline and Dominic Insalaco, whose help and guidance have made me what I am today. I cannot enough express my gratitude.

It is also in part dedicated to the late Dr. Kenneth E. Damann who has left us far too early.

### Biographical Sketch

Born [REDACTED] [REDACTED], Samuel Edward Insalaco grew up in Rochester, New York. He attended [REDACTED] [REDACTED]. After graduation in 1971, he began pursuing a liberal arts degree at the State University of New York at Brockport and transferred shortly after to Monroe Community College. In 1973, he reentered the State University of New York at Brockport and in 1975 was granted a Bachelor of Science degree in biological sciences. He continued on for a Masters of Science degree in Aquatic Ecology and was presented with a Sigma Xi Graduate Research Award in 1978. Upon completion in 1979, he was employed as a Research Scientist with the New York State Health Department, Division of Laboratories and Research, Albany, New York.

Routine Determination of Mirex and Photomirex in Fish  
Tissue in the Presence of Polychlorinated Biphenyls

Samuel E. Insalaco<sup>1</sup> and Joseph C. Makarewicz\*

Department of Biological Sciences

Martha M. Vestling

Department of Chemistry

State University of New York at Brockport

Brockport, New York 14420

Abstract

A procedure for the routine determination of Mirex and Photomirex in fish tissue is described which provides rapid analysis and confirmation using conventional gas chromatographic/electron capture detection (GC/ECD) methods. Coeluting interferences (i.e. PCB's) are nitrated allowing for simple separation from Mirex analogs by column chromatography. In Chinook Salmon tissue (Oncorhynchus tshawytscha), PCB removal averaged 78% and Mirex and Photomirex recoveries were 91% and 86%, respectively. The method has been used successfully for trace analysis of Mirex levels as low as 100 pg.

\* Correspondence

<sup>1</sup> Present address - New York State Department of Health, Division of Laboratories and Research, Empire State Plaza, Albany, New York 12207

Key words: Mirex, Photomirex, fish tissue, analysis

## Introduction

The pesticide chemical Mirex was first discovered in fishes from Lake Ontario by Kaiser (1974). His discovery of Mirex [Dodecachloro-octahydro-1,3,4-metheno-2H-cyclobuta (c,d) pentalene] was based on computerized mass spectral analysis indicating the presence of a m/e 272  $[C_5Cl_6]^+$  ion, an ion not seen in PCB mass spectral fragmentation patterns. Kaiser's work indicated that conventional gas chromatographic/electron capture detection (GC/ECD) analysis of fish samples gave misleading results with respect to trace contaminants because PCB's and Mirex have overlapping retention times. Many laboratories had probably been misinterpreting Mirex as a part of a PCB isomer peak. Similar situations had occurred previously with PCB interference in GC/ECD analyses of dichlorodiphenyltrichloroethane (DDT) and other related chlorinated pesticides (Reynolds 1969, Bonelli 1971, Gustafson 1970).

The discovery of Mirex in Lake Ontario stimulated research by various academic groups and regulatory agencies. Several analytical methodologies for GC/ECD analysis of Mirex were developed; many were modeled on PCB/chlorinated pesticide column chromatographic separation techniques (Reynolds 1969, Armour and Burke 1970, Holden and Marsden 1969). Although high yields and separation efficiencies were reported for these methods, the procedures were not entirely satisfactory. Reproducibility was difficult to maintain in both adsorbent and solvent systems (Berg et al. 1972, Holdrinet 1974, Task Force on Mirex 1977).

Further analytical complications arose with the discovery in fish tissue of a new group of degradation products of Mirex (Hallett et al. 1976). Photomirex [8-monohydromirex], with concentrations as high as 50% of the

reported Mirex values (TFM 1977), also coelutes with PCB's making analysis by conventional GC/ECD techniques difficult.

A simple analytical procedure for separating Mirex and Photomirex from PCB's is needed for rapid routine assessment of the environmental impact of these persistent pesticides. Initially, our objective was to determine the concentrations of Mirex in Lake Ontario fish and to determine where Mirex accumulates within the fish. To accomplish this, we developed and report on a routine analytical procedure for Mirex and Photomirex. Also, our preliminary results on pesticide distribution within fish are presented.

One general approach to the problem of separating Mirex and Photomirex from PCB's is to chemically alter the PCB's by perchlorination or nitration and change their chromatographic behavior. Perchlorination converts all PCB isomers to the decachlorobiphenyl isomer by means of antimony pentachloride. In routine GC analyses, decachlorobiphenyl elutes sufficiently beyond Mirex, thereby yielding adequate separation. The second type, nitration, converts PCB's to nitrated PCB's (hydrogen replacements) with fuming nitric acid. The nitrated PCB's are then separated from Mirex by means of column chromatography.

The perchlorination method was developed by Armour (1973) from work originally presented by Berg *et al.* (1972). Hallett *et al.* (1976) has used a perchlorination technique for the analysis of tissue samples. It is claimed that this method has the strong advantage of PCB removal for Mirex analysis and total PCB quantitation with high recoveries for both ( $99 \pm 0.5\%$  at ppm levels). However, the results of perchlorinating mixtures known to contain Photomirex have not yet been reported. If Photomirex is perchlorinated under the conditions reported, the amounts of Mirex using this method may be misleadingly high.

A nitration procedure for use with tissue samples has been developed by Norstrom et al. (1978). Unlike perchlorination, total PCB's can not be quantitated. However, no confusion concerning chlorination of Photomirex surrounds this method. We have modified Norstrom's (1978) procedures for our use.

### Experimental

Chinook salmon (Oncorhynchus tshawytscha) collected from Lake Ontario were filleted into six predefined tissue sections based on the types and amounts of fat present in the fish. The tissue sections were: (A) the whole fillet; (B) the skin; (C) the "red muscle" (Muscle lateralis superficialis); (D) the belly flap; (E) the anterior dorsal loin; and (F) the caudal peduncle (Fig. 1). A total of 18 tissue samples were analyzed (three samples per tissue section).

Each tissue section was ground and mixed thoroughly in an Intedge food processor. A five-gram aliquot was mixed with 20 g of anhydrous sodium sulfate and ground to a homogenous mixture using a Virtis "45" tissue homogenizer. The contents were packed into an extraction thimble and extracted overnight (a minimum of 200 Soxhlet cycles) in a Soxhlet extraction apparatus with 75 ml of a solvent mixture containing 20% methylene chloride and 80% hexane (20/80 solvent mixture). The 20/80 solvent mixture was used because it gave partitioning coefficients of  $\approx 1.0$  (Galway 1977) as opposed to 0.87 for n-hexane and 0.39 for benzene (Norstrom 1977).

A 15-ml aliquot of the extraction solvent was concentrated to  $\approx 1$  ml in a 15-ml graduated centrifuge tube under nitrogen. The tube walls were then washed with 5 ml of the 20/80 solvent mixture and again concentrated to  $\approx 1$  ml.

The 1-ml sample and two subsequent 1-ml rinses of the tube with the 20/80 solvent mixture were placed on a Florisil column for cleanup.

Florisil columns were constructed using 1-cm (i.d.) burets cut to ~25 cm in length. Columns were plugged initially with glass wool and topped with a thin layer of sand. Five grams of Florisil (60-100 mesh, ranging from  $1\frac{1}{2}$  -  $2\frac{1}{2}$  water by weight) was slurried in 100 ml of 20/80 solvent mixture inside a 500-ml dropping funnel before column packing. This procedure made column packing convenient and hastened adsorbent/solvent equilibration. The slurry was added to a partially filled solvent column (stopcock open) and tapped gently to promote even packing. The column was topped with ~1.5 cm of anhydrous sodium sulfate. This procedure routinely gave a uniform packing.

The 20/80 solvent mixture was used to elute the sample into 50-ml graduated centrifuge tubes. With an elution rate of ~4 ml/min, a total of 50 ml was collected. Depending on the elution rate, the final volume collected should be adjusted for maximum recovery of the pesticide of interest (i.e. Mirex, Photomirex or Mirex and Photomirex together). More information on this is presented in the results section. The eluate was slowly evaporated to dryness under nitrogen, the tube walls washed with 5 ml of the 20/80 solvent mixture, and again evaporated to dryness.

A nitration procedure was used to reduce interfering PCB's. The method is similar to that described by Norstrom (1977) but differs in sample/reagent amounts and the type of adsorbent used for separation. These changes facilitated handling and resulted in increased recoveries. Nitration reagent was prepared from 90% red fuming nitric acid and concentrated sulfuric acid as a 1:1 mixture. Five ml of the nitration reagent was added to the sample tube and the tube was stoppered. After swirling to thoroughly coat the tube walls, the sample tube was placed in a water bath at 70°C for 30 minutes.

The tube was then removed and cooled in ice. Carefully, 10 ml of high purity distilled water was added and the sample vortexed to insure homogeneity. This deliberate reverse order of adding water to an acid reduces sample loss by transfer and is safe if the sample is iced and water is added slowly. The tube was allowed to come to room temperature, and 10 ml of the 20/80 mixture was added. After vortexing for a minimum of 30 seconds, the emulsion was allowed to separate, and exactly 5 ml of the solvent layer was removed into a 15-ml graduated centrifuge tube. The solvent was completely evaporated under nitrogen, and the tube was washed with 5 ml of hexane only. The sample was concentrated to ~1 ml under nitrogen. It is important that all the methylene chloride from the 20/80 solvent mixture be removed prior to Florisil separation. We have found that small amounts of methylene chloride will cause the nitrated PCB's to elute prematurely from the column with the Mirex analogs. A similar situation also occurs with the alumina column separation (Norstrom 1977).

Florisil "micro" columns were prepared to separate the nitrated PCB's from the Mirex analogs. Short-barrelled Pasteur pipets were plugged with glass wool and packed with an unactivated Florisil/hexane slurry to a length of ~5 cm. To insure uniformity we used a 1-ml syringe to pack the Florisil slurry (60-100 mesh). Micro-columns were topped with ~1 cm of anhydrous sodium sulfate and stored in hexane until used.

The 1-ml hexane sample was added to the column followed by two 1-ml hexane "washes" of the tube. A total of 15 to 20 ml of the hexane eluate was collected in a 50-ml graduated centrifuge tube. Again the volume eluted should be adjusted for maximum recovery of the pesticide of interest (i.e. Mirex, Photomirex, or Mirex and Photomirex together). Beyond 20 ml, there

is a strong tendency for the nitrated PCB's to elute from the column. The sample was concentrated under nitrogen to a final volume of exactly 1 ml and is ready for GC analysis. With columns prepared in advance, sample processing time can be shortened to ~4 to 6 hours after extraction.

A Hewlett-Packard 5750-B Research Gas Chromatograph, equipped with a high temperature  $^{63}\text{Ni}$  pulsed electron capture detector, was used to analyze samples. The primary column used was a 4' x 1/8" glass packed with 3.8% UCW-982 on 80/100 Chromosorb W.H.P. An alternate column consisting of 6' x 1/8" stainless steel packed with 4% SE-30 on 60/80 Chromosorb W-AW was also used. Operating parameters were as follows: Pulse interval, 50  $\mu\text{sec}$ ; Carrier gas and Purge gas, 95% argon - 5% methane at 25 and 50 ml/min, respectively; Injection port temperature, 210 $^{\circ}\text{C}$ ; Column temperature, 200 $^{\circ}\text{C}$ ; and Detector temperature, 230 $^{\circ}\text{C}$ . Retention times for Photomirex and Mirex were 23 and 35 minutes, respectively. GC peak areas were determined by polar planimetry. Mirex, Arochlor 1254, and Arochlor 1260 standards were obtained from Chem Services. Photomirex standard was prepared by the Mississippi State Chemical Laboratory, Mississippi State, Mississippi.

### Results

We define nitration efficiency as the reduction in PCB peak size of a sample after nitration expressed as percent of the PCB peak size before nitration. To evaluate the removal of interfering PCB peaks by the nitration procedure, we analyzed salmon tissue and a series of test mixtures containing Mirex, Photomirex, and PCB standards. The two test mixtures established were: Group I (25  $\mu\text{g}$  Arochlor 1254, 25  $\mu\text{g}$  Arochlor 1260, 2  $\mu\text{g}$  Photomirex, and 2  $\mu\text{g}$  Mirex), and Group II (50  $\mu\text{g}$ , 50  $\mu\text{g}$ , 4  $\mu\text{g}$ , and 4  $\mu\text{g}$ ).

For test mixtures I and II, the nitration efficiency averaged  $97\% \pm 0.4$  ( $\bar{x} \pm$  S.E.) indicating a 97% reduction in PCB peak size (Fig. 2).

The effect of the length of the heating period during nitration on the test mixtures was evaluated by adjusting the duration of heating time from 30 minutes to 8 hours and later to 16 hours. No significant changes in peak resolution, retention time, nitration efficiencies, or recoveries of Mirex or Photomirex were evident between test mixtures I and II (Table 1). With the 8 and 16-hour heating times, no nitration of Photomirex occurred. Photomirex is apparently treated as an alkane and does not undergo electrophilic aromatic substitution.

In the salmon tissue, nitration efficiency decreased with samples of high oil content (i.e. skin > belly flap > red muscle). To alleviate this problem, larger amounts of Florisil were initially used for column chromatography. However, the larger amounts of Florisil caused handling problems.

Kates (1972) notes that solvent elution rate is critical for proper column performance. By reducing the amount of Florisil back to the initial 5 g and maintaining the elution rate at  $\sim 4$  ml/min, improved precision and increased nitration efficiencies ( $78 \pm 2.0\%$ ) for these oily samples were achieved (Table 2). Although the nitration was not as efficient as in the test standards, the procedure allowed for more than adequate removal of PCB's in the tissue samples (Fig. 3).

Recovery efficiencies were calculated for the standard test mixture and fish tissue by spiking in the following manner. Standard test mixtures were concentrated under nitrogen in the reaction tubes. These mixtures were nitrated and analyzed as described in the methods section. Using a micro-syringe, we spiked the fish tissues by injecting a standard directly into

the tissue sample while the tissue was in the homogenizer cup. After 5 minutes, the sample was ground and analyzed by the procedure described in the methods section. Rinse blanks of the empty homogenizer cup after grinding indicated that no detectable amounts of the spike were present. The recovery of Mirex and Photomirex added before the initial Florisil column cleanup was  $95 \pm 0.2\%$  ( $\bar{x} \pm$  S.E.);  $n = 9$  for Group I and Group II, total of 18.

Recovery efficiencies for "spiked" tuna fish (packed in oil) in tissue samples were lower than those for standard test mixtures. Mirex averaged  $91 \pm 0.8\%$  as compared to  $95 \pm 0.2\%$  for the test mixtures. Photomirex recoveries were initially low,  $62 \pm 1.3\%$ . This low recovery rate appears to be associated again with the Florisil lipid cleanup. During column cleanup, Mirex repeatedly eluted completely within the first 20 to 30 ml. Photomirex would begin to elute in the 30 to 40 ml-fraction and would sometimes require up to 100 ml of solvent to elute completely. When the elution rate was optimized at 4 ml/min, Photomirex recoveries were increased to  $85 \pm 3.1\%$  and both Mirex and Photomirex were completely eluted within 50 ml (Table 3). In the experimental section, we recommend an elution rate of  $\sim 4$  ml/min.

### Discussion

The method described in this paper provides for the routine tissue analysis of Mirex and Photomirex in the presence of high levels of PCB's. The relative ease of analysis, simplistic design, and quantitation of Mirex and Photomirex without sophisticated instrumentation make this procedure attractive. Although quantitation of PCB's is not possible, the method does offer high sensitivity for Mirex and Photomirex. In our laboratory,

the procedure has been capable of quantitating Mirex at levels as low as 100 pg.

Our preliminary results of Mirex and Photomirex distribution in various tissues of salmon are given in Table 4. Mirex levels are similar in the whole fillet, skin, red muscle, and belly flap. Mirex levels in the anterior dorsal loin and caudal peduncle are comparable but are half the levels seen in the other tissue sections.

Photomirex levels mimic the Mirex distribution in that Photomirex concentrations in the whole fillet, skin, and belly flaps are similar. However, concentrations in the red muscle are low, with levels similar to those seen in the anterior dorsal loin and caudal peduncle. The significance of this result is not fully understood at this time.

Our results suggest that the recommendation made by state and federal agencies (N.Y.S. Department of Environmental Conservation 1978, Duttweiler and Voiland 1979) on dressing of salmonids to reduce Mirex intake in humans consuming fish is correct. However, our results represent only spring fish and are of a preliminary nature.

Much is already known about Mirex's entry and accumulation in the lake, toxicity, and persistency. The occurrence of Photomirex in high concentrations (Hallett et al. 1976, TFM 1977, Table 4) within lake biota is of interest, especially since it is more toxic than Mirex (Hallett et al. 1978, Villeneuve et al. 1979). Since Mirex is known to photolyse to Photomirex in the environment (Carlson et al. 1976, Ivie et al. 1974), it is possible that Photomirex levels may be increasing in the environment. However, no data is currently available to dismiss or substantiate this hypothesis. Furthermore, the chemical pathway and site of conversion of Mirex into Photomirex into the lake ecosystem is poorly understood. Trophic level studies of Mirex and Photomirex in the Lake Ontario ecosystem could provide valuable additional insight into organochlorine pesticide kinetics.

### Acknowledgements

The authors thank W. Abraham of the New York State Department of Environmental Conservation and G. Genga of EcoResearch for their help in obtaining fish specimens for this project. This study would not have been possible without the consultation of T. Haines of the U.S. Fish & Wildlife Service. We thank L. Kline for critical evaluation and discussion. Financial assistance was provided by the Department of Biological Sciences (T. Starr, Chairman), SUNY Brockport.

Table I Effects of heating duration on nitration and recovery efficiencies for Mirex and Photomirex. (combined) between test groups I and II. Three determinations (n=3) were made for each treatment.

	<u>Heating Duration</u>		
	30 min.	8 hours	16 hours
<u>Nitration Efficiency (%)</u>		$\bar{x} \pm \text{S.E.}$	
Group I	96 ± 0	97 ± 0	98 ± 0
Group II	96 ± 0	97 ± 0	98 ± 0
		97 ± 0.4% ( $\bar{x} \pm \text{S.E.}$ )	
<u>Recovery Efficiency (%)</u>		$\bar{x} \pm \text{S.E.}$	
Group I	95 ± 1.0	95 ± 2.1	95 ± 0.6
Group II	94 ± 2.0	95 ± 0.6	95 ± 1.5
		95 ± 0.2% ( $\bar{x} \pm \text{S.E.}$ )	

Table 2 Initial and improved nitration efficiencies from different tissue samples. Three determinations (n=3) were made for each treatment.

<u>Tissue Section</u>	<u>Initial</u> ( $\bar{x} \pm \text{S.E.}$ )	<u>Improved</u> ( $\bar{x} \pm \text{S.E.}$ )
Whole Fillet	58 $\pm$ 0.9%	77 $\pm$ 0.3%
Skin	47 $\pm$ 0.7%	73 $\pm$ 1.2%
Red Muscle ( <u>Muscle lateralis superficialis</u> )	55 $\pm$ 2.2%	74 $\pm$ 0.3%
Belly Flap	53 $\pm$ 1.7%	78 $\pm$ 0.6%
Anterior Dorsal Loin	75 $\pm$ 1.5%	86 $\pm$ 0.7%
Caudal Peduncle	71 $\pm$ 0.3%	82 $\pm$ 0.6%
	<hr/> $\bar{x} \pm \text{S.E.}$ 60 $\pm$ 4.4%	<hr/> 78 $\pm$ 2.0%

Table 3 Initial Photomirex and improved Photomirex and Mirex recovery efficiencies for different tissue samples. Three determinations (n=3) were made for each treatment.

<u>Tissue Section</u>	Initial	Improved	Improved
	<u>Photomirex</u>	<u>Photomirex</u>	<u>Mirex</u>
	( $\bar{x} \pm \text{S.E.}$ )	( $\bar{x} \pm \text{S.E.}$ )	( $\bar{x} \pm \text{S.E.}$ )
Whole Fillet	59 $\pm$ 1.4%	85 $\pm$ 1.2%	91 $\pm$ 0.3%
Skin	66 $\pm$ 1.7%	86 $\pm$ 2.3%	88 $\pm$ 0.9%
Red Muscle			
( <u>Muscle lateralis superficialis</u> )	59 $\pm$ 3.3%	85 $\pm$ 1.8%	93 $\pm$ 1.2%
Belly Flap	64 $\pm$ 2.3%	84 $\pm$ 1.2%	94 $\pm$ 1.4%
Anterior Dorsal Loin	59 $\pm$ 0.9%	86 $\pm$ 0.3%	91 $\pm$ 1.5%
Caudal Peduncle	64 $\pm$ 1.7%	85 $\pm$ 0.6%	91 $\pm$ 0.6%
	$\bar{x} \pm \text{S.E.}$ 62 $\pm$ 1.3%	85 $\pm$ 3.1%	91 $\pm$ 0.8%

Table 4 Preliminary tissue distribution data of Mirex and Photomirex in Chinook Salmon. Three samples were analyzed (two females, one male; all immature).

Length 47.5 ± 0.5 cm ( $\bar{x}$  ± S.D.)

Weight 1.231 ± 0.08 kg ( $\bar{x}$  ± S.D.)

	Corrected	
	Concentrations (mg/kg)	
	Mirex ( $\bar{x}$ ± S.E.)	Photomirex ( $\bar{x}$ ± S.E.)
Whole Fillet	0.22 ± 0.03	0.13 ± 0.01
Skin	0.23 ± 0.04	0.14 ± 0.02
Red Muscle ( <u>Muscle lateralis superficialis</u> )	0.17 ± 0.01	0.05 ± 0.00
Belly Flap	0.22 ± 0.02	0.14 ± 0.03
Anterior Dorsal Loin	0.10 ± 0.01	0.05 ± 0.00
Caudal Peduncle	0.09 ± 0.00	0.05 ± 0.00
	$\bar{x}$ ± S.E.)	
	0.17 ± 0.02	0.09 ± 0.02

### Literature Cited

- Armour, J., Burke, J., J. Assoc. Off. Anal. Chem. 53, 761 (1970).
- Armour, J. A., J. Assoc. Off. Anal. Chem. 56, 987 (1973).
- Berg, O.W., Diosady, P.L., Rees, G.A.V., Bull. Environ. Contam. Toxicol. 7, 338 (1972).
- Bonelli, E.J., Amer. Lab. 3, 27 (1971)
- Calway, P.G., Dept. of Fisheries & the Environment, Toronto, Ontario.  
Personal Communication (1977).
- Carlson, D.A., Konyha, K.D., Wheeler, W.B. Marshall, G.P., Zaylskie, R.G.,  
Science 194, 939 (1976).
- Duttweiler, M. and M.P. Voiland, Fish Contaminants: Minimizing Your Intake.  
New York State Sea Grant. Available from: Sea Grant Extension, Morgan  
III, SUNY at Brockport, Brockport, N.Y. (1979).
- Gustafson, C.G. Environ. Sci. Tech. 4, 814 (1970).
- Hallett, D.J. Norstrom, R.J., Onuska, F.I., Comba, M.E., Sampson, R.,  
J. Agric. Food Chem. 24, 1189 (1976).
- Hallett, D.J., Khera, K.S., Stoltz, D.R., Chu, I., Villeneuve, D.C.,  
Trivett, G., J. Agric. Food Chem. 26, 388 (1978).
- Holden, A.V., Marsden, K., J. Chromatog. 44, 481 (1969).
- Holdrinet, M.V.H., J. Assoc. Off. Anal. Chem. 57, 580 (1974).
- Ivie, G.W., Gibson, J.R., Bryant, H.E., Begin, J.J., Barnett, J.R.,  
Dorough, H.W., J. Agric. Food Chem. 22, 646 (1974).
- Kaiser, K.L.E., Science 185, 523 (1974).
- Kates, M. Techniques of Lipidology, Amsterdam: North Holland Publishing  
Co., p. 393-404 (1972).

New York State Department of Environmental Conservation. Reducing Toxics: Fish Filleting Guide. Available from: Sea Grant Extension, Morgan III, SUNY at Brockport, Brockport, N.Y. (1978).

Norstrom, R.J., H. Won, M.V.H. Holdrinet, P.G. Calway, and P. Robinson, International Symposium on the Analysis of Hydrocarbons and Halogenated Hydrocarbons, May 23-25, McMaster University, Hamilton, Ontario, Canada (1978).

Norstrom, R.J., National Wildlife Research Centre, Dept. of Fisheries and Environment, Ottawa, Ontario. Personal Communication (1977).

Reynolds, L.M., Bull. Environ. Contam. Toxicol. 4, 128 (1969).

Task Force on Mirex (TFM). A report of the Task Force to the Environmental Contaminants Committee of Fisheries & Environment Canada and Health & Welfare Canada. Technical Report 77-1, 153 p., Ottawa (1977).

Villeneuve, D.C., Ritter, L., Felsky, G., Norstrom, R.J., Marino, I.A., Valli, V.E., Chu, I., Becking, G.C., Toxicol. Appl. Pharmacol. 47, 105 (1979).

## Captions for Figures

- Figure 1 Tissue sections analyzed for Mirex and Photomirex: (A) the whole fillet; (B) the skin only; (C) the "red muscle" (Muscle lateralis superficialis); (D) the belly flap; (E) the anterior dorsal loin; and (F) the caudal peduncle.
- Figure 2 Gas chromatograms of standard test mixture (Group II) before (A - 2  $\mu$ l injection) and after (B - 10  $\mu$ l injection) nitration. 1 and 2 indicate the Photomirex and Mirex peaks, respectively.
- Figure 3 Gas chromatograms of salmon tissue (1 g) with Florisil split (A - 2  $\mu$ l injection) and with Florisil split plus nitration (B - 15  $\mu$ l injection). 1 and 2 indicate the Photomirex and Mirex peaks, respectively.

Figure 1




Figure 2





Figure 3

